भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान मोहाली

शिक्षा मंत्रालय, भारत सरकार द्वारा स्थापित

सैक्टर 81, नॉलेज सिटी, प० ओ० मनौली, एस० ए० एस० नगर, मोहाली, पंजाब - 140306

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH MOHALI

(Estd. By Ministry of Education, Govt of India)

Sector-81, Knowledge city, PO-Manauli, SAS Nagar Mohali-140306, Punjab

PAN No. - AAAAI1781K GST No. 03AAAAI1781K2ZS

● Phone : +91-172-2240121 ● Fax : +91-172-2240124, 2240266 ● http://www.iisermohali.ac.in ● Email: tenders@iisermohali.ac.in

CPPP/Institute Website

IISERM(1699)25/26-Pur-GTE Dated: 07-07-2025

शुद्धिपत्र -I / CORRIGENDUM –I

Refer to IISER Mohali Global Tender Enquiry ref. no. IISERM(1699)25/26-Pur-GTE dated 17-06-2025 for inviting bids towards "**Supply, Installation and Commissioning of Single Crystal X Ray Diffractometer along with Accessories",** Knowledge City, PO. Manauli S.A.S. Nagar, Mohali, Punjab-140306. The On-line bid submission date is further extended due to amendments / additions in NIT in the above GTE notice, **as per below schedule**:

BOQ no.	Details of Specifications of the Stores	Addition / Amendment in technical specifications
1.	**Supply, Installation and Commissioning of Single Crystal X Ray Diffractometer along with Accessories**	**As per NIT / BOQ and as per below amendments**
Last para of Sr. No. 2	Each tube supplied with the instrument should have a minimum of 3-year warranty **covering the generator, HV-cable, tube shield, X-ray tube and the optics. The X-ray generator must be in-house designed and manufactured by the SCXRD manufacturer and should be repairable on component level at site.**	Each tube supplied with the instrument should have a minimum of **5-year warranty covering the generator, HV-cable, tube shield, X-ray tube and the optics. The X-ray generator must be in-house designed and manufactured by the SCXRD manufacturer and should be repairable on component level at site.**
Sr. No. 18	The system should be quoted with a minimum 5 years of comprehensive on-site warranty after completion of warranty period. During warranty period all parts, spare parts, and labor, free maintenance and service should be provided.	The system should be quoted with a minimum 5 years of on-site warranty of all items (X-ray source, No. 2; Detector, No. 3; Goniometer, No. 4; Low temperature set up units, No. 5). During the warranty period all parts, spare parts, and labor, free maintenance and service should be provided.
Optional Item No. 2	AMC services for 5 years after the expiry of 3 years warranty and 5 years CMC (point #18 above). This should include minimum 2 preventive maintenance and 1 emergency visit. Vendor has to include all Preventive maintenance parts required during each AMC service.	AMC services for 5 years after the expiry of 5 years warranty (point #18 above). This should include minimum 2 preventive maintenance and 1 emergency visit. Vendor has to include all Preventive maintenance parts required during each AMC service.

Extension of end time for submission of Bid

*Bidders are requested to furnish bid incorporating the above changes/ amendment/s.

Revised date and Time:

Sr. No.	Description	Extended Due date/Time
01	**Closing Date & Time (Online) submission**	**21st July 2025 upto 2:30 PM**
02	**Opening Date & Time of Technical Bid**	**22nd July 2025 upto 3:30 PM**

For any information, other modifications and corrigendum etc may kindly visit IISER Mohali websites http://www.iisermohali.ac.in/ & CPPP Portal https:/eprocure.gov.in/eprocure/app

Sd/-

सहायक कुलसचिव (क्रय तथा भंडार)

Assistant Registrar (Stores & Purchase)