Original Tender Specifications	Revised/ Updated Tender Specifications
[A] Mandatory Specifications: (4) Optical System & Objectives: Infinity corrected POL; parfocal, centrable; 2.5× (opt), 5×, 10×, 20×;	**[A] Mandatory Specifications:** 4) Infinity-corrected, strain-free POL objectives; parfocal and centrable; objective magnifications: 2.5× (or 2.0 x), 5×, 10×, 20×; all objectives should be of Semi-Apochromat / Plan Fluotar / FL Semi-Apo / Plan-Neofluar / Plan Fluor grade or equivalent fluorite-corrected class, suitable for polarizing microscopy and fluid/mineral inclusion observations.
[B] Optional Requirements: (1) Objectives: 50X;	**[B] Optional Requirements:** 1) 50× strain-free, infinity-corrected POL objective; parfocal and centrable; dry or oil immersion; Semi-Apochromat / Plan Fluotar / FL Semi-Apo / Plan-Neofluar / Plan Fluor grade or equivalent fluorite-corrected objective.