Corrigendum:

<table>
<tr><td colspan="2">Original Tender Specifications</td><td>To be read as</td></tr>
<tr><td>Human Machine Interface (HMI: 7-inch)</td><td>Graphical display for all three reactors: pH, temperature, and agitation speed</td><td>Graphical display for all two reactors: pH, temperature, and agitation speed</td></tr>
<tr><td colspan="3">Further clarifications:
1. Separate individual control panels are required for each fermenter. Accordingly, individual control panels should be provided for each fermenter.
2. Utilities such as air, steam, water, etc., shall be provided as combined/common utilities for both fermenters. Thus, separate utility systems for each fermenter are not required.
3. Automation is required for each individual fermenter. Any suitable automation solution that enables independent monitoring, control, and data logging for each reactor is acceptable.</td></tr>
</table>