भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान मोहाली

मानव संसाधन एंव विकास मंत्रालय, भारत सरकार द्वारा स्थापित
सैक्टर 81, नॉलेज सिटी, पी. ओ. मनोली, एस. ए. एस. नगर, मोहाली, पंजाब –140306
INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH MOHALI
Sector-81, Knowledge city, PO-Manauli, SAS Nagar Mohali-140306, Punjab
PAN NO. - AAAAI1781K TAN NO. PTLI10692D

IISER Mohali

● Phone : +91-172-2240086,2240121 ● Fax : +91-172-2240124, 2240266 ● http://www.iisermohali.ac.in ● Email: stores@iisermohali.ac.in

E-mail/CPPP /website

IISERM(631)15/16 Pur 15th February 2016

NOTICE INVITING QUOTATION

Sealed Quotations are invited on behalf of The Director IISER Mohali for following items along with EMD money by DD /Banker Cheque/ FDR/ Bank guarantee for Rs 5000/- and tender fee of Rs 500/-(separately non re-fundable) in favour of the Registrar, IISER Mohali, payable at Mohali/Chandigarh so as to reach us latest by ***08.03.2016*** before 1pm and will be opened on the same day at 4pm. in the presence of tenderers, if any :-

Sr.	Description	Qty
01	**5-digit analytical balance (with windows)** Should work both as 4 digit and 5 digit analytical balance Single block weighing technology , Capacity :minimum 200g for 4 digit / minimum 50g for 5 digit , Readability: 0.1 for 4 digit / 0.01 mg for 5 digit Pan size (diameter) 80 mm or higher., Repeatability : 0.1 mg for 4 digit / 0.05 mg for 5 digit , Linearity: ± 0.2 mg for 4 digit / ± 0.1 mg for 5 digit Response time: of the order of few seconds (please mention) Operating Temperature : upto 40℃ , Option for hanging measurement. Fully automatic calibration, Fully automatic calibration at user pre set time. Backlit display , Balance should send data to the PC in excel or other window applications at set time intervals. Any additional software required for this should be quoted.	01
02	**4-digit precision balance (with windows)** Single block weighing technology. , Pan Size (diameter): 90 mm or more. Capacity : 200 g or more , Readability: 0.1 mg , Repeatability (Standard Deviation) : ≤0.1mg , Linearity: ± 0.2 mg Stabilisation Time: 2-3 seconds , Operating Temperature : upto 40℃ Automated internal calibration. , Multiple weighing units. Balance should send data to the PC in excel or other window applications at set time intervals. Any additional software required for this should be quoted	01
03	Balance should have Power Saving Function i.e. **2-digit precision balance (with windows)** Weighing capacity : approximate 600g. , Readability : 0.01g. Pan Size (diameter): 100 mm or more , Repeatability : 0.01g Linearity : +/- 0.02 gms , External calibration provision. Backlit display. , Power Saving Function , Provision for In-built battery back-up. Balance should send data to the PC in excel or other window applications at set time intervals. Any additional software required for this should be quoted.	01

NB:- Suitable UPS back up for the quoted balances may also be provided.

-sd-
(Mukesh Kumar)
Assistant Registrar (S&P)

INSTRUCTIONS

1. The Quotation should be addressed to the Assistant Registrar (S&P), IISER Mohali invariably giving on the envelop Reference number with due date and time of submission of quotation.

2. One time importers from China with custom made specifications are highly discouraged.

3. The quantity mentioned in this inquiry is and shall be deemed to be only approximate and will not in any manner be binding on the Institute.

4. The rates offered should be FOR Chandigarh/Mohali in case of firms situated outside Chandigarh/Mohali, and free delivery at the Institute premises in case of local firms. Supplier outside India please mentions the Ex-works/FOB/FCA/CIF/CIP price clearly.

5. In case of Ex-godown terms the amount of packaging forwarding freight etc should clearly be indicated by percentage or lump sum amount. Institute has policy not to make any advance payments towards any purchase, Letter of credit can be opened if required.

6. THE INSTITUTE IS EXEMPTED FROM EXCISE AND CUSTOM DUTY.

7. SALES TAX :- This Institute is not exempted from the payment of Sales Tax/Service Tax/VAT. The current rate (i.e. percentage of Sales Tax should be clearly indicated included or excluded) wherever chargeable.

8. The delivery period should be specifically stated. Earlier delivery will be preferred

9. The firms are requested to give detailed description and specifications together with the detailed drawings, printed leaflets and literature of the article quoted. The name of the manufactures and country of manufacture should also invariably be stated. In the absence of these particulars, the quotation is liable for rejection.

10. Validity of offer: 90 days. Firm should replace all manufacturing defect parts/ whole item under warranty period without any extra cost including clearance, freight, taxes. Security deposit/ Bank Performance guarantee for 10 % of the value of supply order as per norms may be sought from firms.

11. Late or delayed quotation will not be accepted.

12. The right to reject all or any of the quotation and to split up the requirements or relax any or all the above conditions without assigning any reason is reserved. For any corrigendum and addendum please regularly visit our website.

.