



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान मोहाली

मानव संसाधन विकास मंत्रालय, भारत सरकार द्वारा स्थापित
सैक्टर 81, नॉलेज सिटी, पं० ओ० मनोली, एस० ए० एस० नगर, मोहाली, पंजाब 140306
INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH MOHALI

(Established by Ministry of Human Resource Development, Govt. of India)
Sector-81, Knowledge city, PO-Manauli, SAS Nagar Mohali-140306, Punjab
PAN No. - AAAAI1781K

• Phone : +91-172-2240086 & 2240121 • Fax : +91-172-2240124, 2240266 • <http://www.iisermohali.ac.in> • Email: stores@iisermohali.ac.in

IISERM (921)17/18Pur.

5th December 2018

Corrigendum-II

Refer IISER Mohali E-tender Ref. no. IISERM (921)17/18 Pur for **SUPPLY AND INSTALLATION OF FUME HOODS AND LAB FURNITURE** at IISER Mohali, Sector-81, Knowledge City, PO-Manauli, SAS Nagar Mohali, Punjab. Due date for the online submission is extended due to **Amendments in the Technical Specifications mentioned below** as per below schedule.

A) AMENDMENT IN TECHNICAL SPECIFICATIONS

Item No. 1: Fume Hood 1800 mm

Section 2.3. Specifications

Overall Fume Hood Structure: 2.0 mm thick Galvanized Iron (GI) powder coated.

External Panels: 1.2mm thick Galvanized Iron steel, coated with epoxy powder with at least 50-micron thickness for high chemical resistance.

Inner Panels & Baffles: At least 6mm thick panels with excellent chemical & heat resistance and fire retardant. These panels should have a smooth finish and easy clean ability.

Item No. 3 to 20: Lab Furniture

Point 1. SEFA-8M (2016 Standards)- the vendor (s) should provide at least two standards out of the stipulated ones.

Point 2. The laboratory furniture should be made of Galvanized Iron duly coated with at least 50 micron Epoxy Powder.

Point 3. Thickness of Galvanized Iron used in construction of cases shall be at least 1 mm, except the following:

- Case and drawer suspension channels, 2 mm thick.
- Top and intermediate front horizontal rails, table aprons, hinge reinforcements, and reinforcement gussets, 1.6 mm thick.
- Drawer assemblies, door assemblies, bottom, bottom back rail, toe space rail, and adjustable shelves are at least 1 mm thick.
- Drawer and Door front panels are in double skin construction-at least 0.8 mm thick.

Item No. 21: Drying Shelves above wall sinks of size 600 x 900 mm: Required Quantity is 5 Nos. instead of 2 Nos.

Item No. 22: Big sink unit in one piece, moulded in Polypropylene construction of overall size 1500x900x900mm with granite top (BSU): 1 Nos.

Item No. 23: Wall side sink unit in one piece, moulded in Polypropylene construction of overall size

600x750x900mm with granite top (WSU): Required Quantity is 4 Nos instead of 1 Nos.

Note No. 7: Warranty for 2 years

Note No. 8: Thickness of granite should be as 18 mm

B) Revised Date and Time

Sr.no.	Description	Extended Due date/Time
01	Closing Date & Time (Online)	18/01/2018 up to 11:00 AM
02	Opening Date & Time of Technical Bid	19/01/2018 at 11:30 AM

All the other terms & conditions will remain same as contained in the NIT.

For any information, other modifications and/or corrigendum may kindly visit IISER Mohali websites <http://www.iisermohali.ac.in> & <https://eprocure.gov.in/eprocure/app>

Sd/-
Mukesh Kumar
Assistant Registrar (S&P)