



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान मोहाली

मानव संसाधन एवं विकास मंत्रालय, भारत सरकार द्वारा स्थापित
सैक्टर 81, नॉलेज सिटी, पी. ओ. मनोली, एस. ए. एस. नगर, मोहाली, पंजाब 140306

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH MOHALI

(Established by Ministry of Human Resource Development, Govt. of India)

Sector-81, Knowledge city, PO-Manauli, SAS Nagar Mohali-140306, Punjab

PAN No. - AAAAI1781K TAN No. PTLI10692D

• Phone : +91-172-2293118-19 • Fax : +91-172-2240124, 2240266 • <http://www.iisermohali.ac.in> • Email: stores@iisermohali.ac.in

E-mail/CPPP/Website

IISER M(452)14/15 Pur.

Date : 09th December 2014

NOTICE INVITING QUOTATION

Dear Sir,

Sealed Quotations are invited on behalf of Director IISER Mohali for the purchase of following items so as to reach latest by 31st December 2014 before 1 pm. The Quotations will be opened on the same day at 4 pm in the presence of tenderers, if any :-

Sr.	Description	Qty
01.	Floor Model Incubator Shaker (Specifications attached)	01
02.	Multipurpose Benchtop Centrifuge (Specifications attached)	01

Yours faithfully,

Sd/-

(Mukesh Kumar)

Assistant Registrar (S&P)

INSTRUCTIONS

1. Inquiry will be sent by UPC/Courier/Speed Post and IISER Mohali will not be liable for any kind of Postal delay.
2. The Quotation should be addressed to the Director IISER Mohali invariably giving on the envelop Reference No. with date and Due date with time.
3. One time importers from China with custom made specifications are highly discouraged.
4. The quantity mentioned in this inquiry is and shall be deemed to be only approximate and will not in any manner be binding on the Institute.
5. Firms will quote separately for each article.
6. The rates offered should be FOR Chandigarh/Mohali in case of firms situated outside Chandigarh/Mohali, and free delivery at the Institute premises in case of local firms. Supply outside India please mention the FOB/FCA price clearly.
7. In case of Ex-godown terms the amount of packaging forwarding freight etc should clearly be indicated by percentage or lump sum amount. Institute has policy not to make any advance payment towards any purchase, Letter of credit can be open if required.
8. In case of foreign consignment CIF/FOB should be mention clearly. THE INSTITUTE IS EXEMPTED FROM EXCISE AND CUSTOM DUTY.
9. SALES TAX :- This Institute is not exempted from the payment of Sales Tax/Service Tax/VAT. The current rate (i.e. percentage of Sales Tax should be clearly indicated included or excluded) wherever chargeable.
10. The delivery period should be strictly adhered, the firm have to be execute the supply according to purchase order.
11. The firms are requested to give detailed description and specifications together with the detailed drawings, printed leaflets and literature of the Article quoted. The name of the manufactures and country of manufacture should also invariably be stated in the absence of these particulars the quotation is liable for rejection.
12. Validity of offer: 90 days.
13. Late or delayed quotation will not be accepted.
14. Service Facility: Please mention the nearest service centre to IISER Mohali and the time required to attend the call
15. The right to reject all or any of the quotation and to split up the requirements or relax any or all the above conditions without assigning any reason is reserved.

Specifications :

Sr.	Items	Description
01.	Floor Model Incubator Shaker	<ol style="list-style-type: none"> 1. Vendors should quote one floor-model, heavy-duty, incubator-shaker suitable to run in year-round weather conditions in non-temperature-controlled laboratory in Chandigarh-Mohali region in India. 2. The range of shaking-speed should be between 50 to 400 rpm with +/- 1 rpm. 3. The shaker should be programmable for shaking period from 0.1 hrs to 99.9 hrs. 4. The shaker should be able to run at temperature 15 degree below ambient (Minimum set point is 4 degree) to 60 degree with in +/- 0.1 deg C and uniformity +/- 0.25 deg C. 5. Temperature and speed should be calibratable through key pad (integrated with shaker). 6. Should start automatically after power interruption with non-volatile memory. 7. Should give audio and visible alarms whenever speed deviates more than 5 rpm and temperature 1 degree centigrade. 8. Universal platform should have capability to hold clamps for 50ml to 6L flasks. Universal platform size should be at least 76 cm X 46 cm. 9. The shaker should have eccentric counter balance drive. 10. Power consumption should not be more than 30 units/day. 11. Refrigeration in the shaker should be designed such that no condensation or frosting is formed during operation. 12. Shaker should be supplied with one microtitre plate holder, which can stack at least 9 microtitre plates in one. 13. Shaker should be supplied with one test-tube rack mountable on the platform. 14. The shaker should be supplied with clamps to hold:500ml flasks-6nos., 1L flasks-6nos., 2L flasks-6nos., and 4L flasks -6nos. 15. Vendors should give 3 years warranty. 16. Vendors should give quote previously quoted models in India. 17. Vendors should provide name and address of at least one research Institution in India where the unit is being used successfully for at least the past one year.
02.	Multipurpose Benchtop Centrifuge	Centrifuge with capacity for (30x1.5ml), (6x15ml), (6x50ml) Along with adapters required. Should have capacity to go upto 30,000g