



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान मोहाली

मानव संसाधन एवं विकास मंत्रालय, भारत सरकार द्वारा स्थापित
सैक्टर 81, कॉलेज सिटी, पी. ओ. मनोली, एस. ए. एस. नगर, मोहाली, पंजाब 140306

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH MOHALI

(Established by Ministry of Human Resource Development, Govt. of India)

Sector-81, Knowledge city, PO-Manauli, SAS Nagar Mohali-140306, Punjab

PAN No. - AAAAI1781K TAN No. PTLI10692D

• Phone: +91-172-2293118-19 • Fax: +91-172-2240124, 2240266 • <http://www.iisermohali.ac.in> • Email: stores@iisermohali.ac.in

E-mail/CPMP/Website

IISERM (538) 15/16 Pur.

Date: 12th May 2015

NOTICE INVITING QUOTATION

Sealed Quotations in **Single bid system** are invited on behalf of Director IISER Mohali for the purchase of following item(s) so as to reach latest by 02nd June 2015 before 1 pm. The Quotations will be opened on the same day at 4 pm in the presence of tenderers, if any

Sr.	Descriptions	Quantity
1.	Shaking Water Bath Specifications:- <ol style="list-style-type: none">1. The water bath should have powerful drive mechanism for quiet operation and smooth consistent shaking over a wide speed range.2. Tank size of the water bath should be at least 20 liters with working volume not less than 15 liters.3. It should have precision digital temperature control with 0-99°C operating range and stability and uniformity of $\pm 0.1^\circ\text{C}$.4. Shaking speed range of the system should be 20 to 200 orbital rpm with orbital shaking of 9 radius rpm.5. Linear shaking range should be from 40 to 300 strokes/min with 3 variable strokes lengths.6. The system should be easily changeable from linear to orbital shaking mode.7. The display for temperature should have resolution 0.1°C and shaking speed 1 strokes/min.8. Heater and temperature sensor should be mounted under tank for large available working area and easy to clean operation.9. System should be upgradable for operation below ambient / to 0°C with accessory cooling coil.10. It should have safety for over-temperature and low-liquid level.11. Vendors should also quote for accessories: Universal Flask Tray, Flat predrilled tray with flask clamp 125 ml – 4 Nos., 250 ml – 4 Nos. and 500 ml – 4 Nos.12. System should be CE Certified and EMC (Electromagnetic Compatibility) Directive 2004/108/EC compatible.13. Minimum 3 years warranty should be included with the system.	01

Sd/-
(Mukesh Kumar)
Assistant Registrar (S&P)

INSTRUCTIONS

1. Inquiry will be sent by UPC/Courier/Speed Post and IISER Mohali will not be liable for any kind of postal delay.
2. The Quotation should be addressed to the Director IISER Mohali invariably giving on the envelop reference number, last date & time of receipt of tender and date & time of opening of the tender.
3. One time importers from China with custom made specifications are highly discouraged.
4. The quantity mentioned in this inquiry is and shall be deemed to be only approximate and will not in any manner be binding on the Institute.
5. Firms will quote separately for each article.
6. The rates offered should be FOR Chandigarh/Mohali, in case of firms situated outside Chandigarh/Mohali, and free delivery at the Institute premises in case of local firms. Suppliers outside India may please mention the FOB/FCA price clearly.
7. In case of Ex-godown terms the amount of packaging forwarding freight etc. should clearly be indicated by percentage or lump sum amount. Institute has policy not to make any advance payment towards any purchase, Letter of credit can be opened if required.
8. In case of foreign consignment CIF/FOB should be mentioned clearly. THE INSTITUTE IS EXEMPTED FROM EXCISE AND CUSTOM DUTY.
9. SALES TAX: - This Institute is not exempted from the payment of Sales Tax/Service Tax/VAT. The current rate (i.e. percentage of Sales Tax should be clearly indicated included or excluded) wherever chargeable.
10. The delivery period should be strictly adhered, the firm have to execute the supply according to purchase order.
11. The firms are requested to give detailed description and specifications together with the detailed drawings, printed leaflets and literature of the Article quoted. The name of the manufactures and country of manufacture should also invariably be stated. In the absence of these particulars, the quotation is liable for rejection.
12. Validity of offer: 90 days.
13. Late or delayed quotation will not be accepted.
14. Service Facility: Please mention the nearest service center to IISER Mohali and the time required to attend the call.
15. The right to reject all or any of the quotation and to split up the requirements or relax any or all the above conditions without assigning any reason is reserved.
16. For any query, tenderers may contact via email : stores@iisermohali.ac.in