

(Technical Compliance Sheet)		Tender
for high frequency signal generator with spectrum analyzer		
Sr.	Description / Specifications	Remarks of Vendor
1	High frequency Spectrum Analyzer specifications	
1.a	Frequency range from 10Hz to 40GHz	
1.b	0.4dB level measurement uncertainty	
1.c	Signal analysis bandwidth of upto 160MHz	
1.d	Including a software option of phase noise measurements	
1.e	Including a software option of electromagnetic interference (EMI) measurement application	
1.f	Should include a removable hard drive	
2	High frequency Signal Generator specifications:	
2.a	Frequency range from 100KHz to 40GHz	
2.b	Sweep for complete frequency range with AM/FM/PM modulations	
2.c	Low signal sideband phase noise < -50 dBc at 40GHz	
2.d	Optional reference oscillator for improved phase noise requirements	

Terms and Conditions :-

1	Price - (FOR/FOB/FCA/CIP/CIF-only)	
2	Validity - (Minimum 90 days)	
3	Delivery Period -	
4	Payment terms - (Letter of Credit / Within 30 days after delivery)	
5	Warranty	

NB :- Please fill in yellow cell only