



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान मोहाली
मानव संसाधन विकास मंत्रालय, भारत सरकार द्वारा स्थापित
सैक्टर 81, नॉलेज सिटी, प० ओ० मनोली, एस० ए० एस० नगर, मोहाली, पंजाब 140306
INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH MOHALI
(Established by Ministry of Human Resource Development, Govt. of India)
Sector-81, Knowledge city, PO-Manauli, SAS Nagar Mohali-140306, Punjab
PAN No. - AAAAI1781K GST No. 03AAAAI1781K1ZT

• Phone : +91-172-2240086 & 2240121 • Fax : +91-172-2240124, 2240266 • <http://www.iisermohali.ac.in> • Email: stores@iisermohali.ac.in

CPPP/Institute Website

IISERM (819)17/18-Pur

Dated :23.08.2017

NOTICE INVITING E-TENDER

Online tenders are invited on behalf of Director, IISER Mohali in **SINGLE BID SYSTEM** for the **Supply & Installation of Ultra Low Temperature (-80) Upright Freezer with Urethane Foam Insulation, 230V, 50Hz and Upright Single Solid Door -20 Degree Freezer with Digital Display & Controller** per technical specification given below and BOQ list the original manufacturer/supplier at CPPP i.e. <https://eprocure.gov.in/eprocure/app>. Tender documents may please be downloaded from the E-procurement portal website <https://eprocure.gov.in/eprocure/app> & Institute website www.iisermohali.ac.in.

-sd-

(Mukesh Kumar)
Assistant Registrar (S&P)



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान मोहाली
मानव संसाधन विकास मंत्रालय, भारत सरकार द्वारा स्थापित
सैक्टर 81, नॉलेजसिटी, प०ओ० मनोली, एस० ए० एस० नगर, मोहाली, पंजाब 140306
INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH MOHALI
(Established by Ministry of Human Resource Development, Govt. of India)
Sector-81, Knowledge city, PO-Manauli, SAS Nagar Mohali-140306, Punjab
PAN No. - AAAAI1781K GST No. 03AAAAI1781K1ZT

• Phone : +91-172-2240086 & 2240121 • Fax : +91-172-2240124, 2240266 • <http://www.iisermohali.ac.in> • Email: stores@iisermohali.ac.in

CPPP/Institute Website

E-TENDER NOTICE

Tender Ref.- IISERM(819)17/18-Pur	Dated :- 23 rd August 2017
-----------------------------------	---------------------------------------

Critical Date Sections

Sr.	Description	Date	Time
1.	Tender Publishing Date and time	23 rd August 2017	6:00pm
2.	Tender Document download start Date & Time	23 rd August 2017	6:00pm
3.	Bid Submission start Date & Time	23 rd August 2017	6:00pm
4.	Bid Submission End date and Time	12 th Sept 2017	Up to 11:00am
5.	Tender opening Date and Time	13 th Sept 2017	At 11.30 am

Online tenders are invited on behalf of the Director, IISER Mohali in **SINGLE BID SYSTEM** for following item(s) from the original manufacturer/supplier at CPPP i.e. <https://eprocure.gov.in/eprocure/app>. Tender documents may please be downloaded from the E-procurement portal website <https://eprocure.gov.in/eprocure/app> & Institute website www.iisermohali.ac.in. Tender fee in shape of DD/Banker Cheque of Rs 500/- (Non-refundable) and EMD of Rs.20000/- should be submitted by DD /Banker Cheque/FDR/ Bank Guarantee in favour of the Registrar, IISER Mohali payable at Mohali. However, scanned copy of the both tender fee and EMD should be upload on website along with technical bid part. The hard copy of the same in original to be send to the address mentioned below duly superscribing the supply/work name and reference/ tender ID on the envelope and same must reach before opening the bid and if not received within due date the bid will be rejected summarily.

The Original EMD and tender fee should be sent to:

Assistant Registrar (S&P)
Indian Institute of Science Education and Research,
Mohali Sector 81, SAS Nagar, Mohali, Punjab, India,
Pin: 140306

Non-receipt of original EMD and tender fee will lead to rejection of tender.

Item Details:

Sr. No.	Details of Specifications of the Stores	Qty.
1.	<p data-bbox="252 277 1193 349">Ultra Low Temperature (-80) Upright Freezer with Urethane Foam Insulation, 230V, 50Hz</p> <p data-bbox="252 387 587 421"><u>Technical Specifications</u></p> <p data-bbox="252 423 740 456">Capacity : 560-600 L; 20-25 Cu. Ft.</p> <p data-bbox="252 459 1286 530">Temp Range: - Programmable temperature range from -50°C to -86°C in 1°C increments;</p> <p data-bbox="252 533 1286 604">Control: Micro Processor control of temperature and alarms with non-volatile memory;</p> <p data-bbox="252 607 1219 640">Circulation: Minimum one fan for reduced power consumption and noise;</p> <p data-bbox="252 642 1286 714">Cascade Refrigeration: Hermetically-sealed two stage cascade system with capacity to cope in high-ambient conditions;</p> <p data-bbox="252 716 667 750">Insulation: Polyurethane foam;</p> <p data-bbox="252 752 679 786">Chamber: Latchable inner door;</p> <p data-bbox="252 788 1286 898">Shelves: Minimum 4 compartment with adjustable height corrosion resistant Stainless Steel Shelves. Each of the compartment with an inner door and Magnetic closures to provide easy access & comfort of use;</p> <p data-bbox="252 900 1286 972">Security: Keyed locks on the outer doors and lids to keep out unauthorized users;</p> <p data-bbox="252 974 1286 1046">Power: On-Off switch behind the locked panel to prevent power from being accidentally turned off;</p> <p data-bbox="252 1048 1273 1081">Battery Backup: Automatic alarms and display temperature for power outage;</p> <p data-bbox="252 1084 1286 1193">Alarms: Audible and Visible alarms for Temperature filter clean, power out, Low Battery, System Fail, and Fault analysis. In-built front control panel to assist trouble shooting fault condition;</p> <p data-bbox="252 1196 443 1229">Construction:</p> <p data-bbox="252 1232 1114 1265">Exterior: Steel with powder coated paint to resist scratch and rust;</p> <p data-bbox="252 1267 501 1301">Interior: Polished;</p> <p data-bbox="252 1303 1286 1375">Door seal: Inner and outer door with low temperature safe silicone seal to prevent temperature loss;</p> <p data-bbox="252 1377 938 1411">Heated Air vent: Automatic vent port on front door;</p> <p data-bbox="252 1413 775 1447">Filter: Front mounted compressor filter;</p> <p data-bbox="252 1449 1286 1520">Electric Power: Programmed startup to prevent power supply overload following a power failure;</p> <p data-bbox="252 1523 683 1556">Stability: with power stabilizers;</p> <p data-bbox="252 1559 737 1592">Certification: CE and CSA certified;</p> <p data-bbox="252 1594 1286 1666">Ambient Condition: Engineered to maintain an internal Temperature of -86°C, even when operated in ambient temperature condition of up to 32-40°C;</p> <p data-bbox="252 1668 1286 1778">Others : System should be installed / validated at the time of installation by Manufacturer / Authorized dealer by their company trained engineer (Latest training certificate should be enclosed.); No third party will be entertained.</p> <p data-bbox="252 1780 1286 1852">No applications will be considered for evaluation if no printed brochure is supplied. Brochure should contain the performance-check results.</p> <p data-bbox="252 1854 1286 1964">Vender should provide the certificate, which they have issued earlier to University/ Government institution/ Private sector in last one year along with atleast 3 user recommendations.</p>	01 No.

2.	<p>Upright Single Solid Door -20 Degree Freezer with Digital Display & Controller</p> <p>Technical Specifications:</p> <ol style="list-style-type: none"> 1) Capacity: 300-350 litres. 2) Dimension: W x D x H (~24 x ~24 x ~73 Inches). 3) Temperature range: -16°C to -25°C. 4) Minimum number of pull-out separate drawers: 5 5) Minimum number of shelves: 3 6) CFC Free Refrigerant and inner body made up of rust free material. 7) System should be provided with digital temperature display and controller. 8) System should have inside light and door lock feature. 9) System should have low power consumption rating. 10) System should be CE certified. 11) System should have wheels for easy movement. 12) Comprehensive warranty: 2 years 13) System should be provided with 1 KVA voltage stabilizer with low and high cutoff and time delay feature. 14) Local service back-up should be provided in case of equipment failure. 15) The system should be operated at 220V and 50Hz. <p>General conditions:</p> <ol style="list-style-type: none"> 1) System manual with the specified information highlighted should be provided with the quotation. 2) Any additional accessory required with the system should be mentioned clearly and quoted separately. 	02 Nos.
----	---	---------

SUBMISSION OF TENDER

- I. All bid/ tender documents are to be uploaded online at Central Public Procurement portal i.e. <https://eprocure.gov.in/eprocure/app> only and in the designated cover/ part on the website against tender ID. Tenders/ bids shall be accepted only through online mode and no manual submission of the same shall be entertained except tender fee and EMD. Late tenders will not be accepted.
- II. The online bids shall be opened at the office of the Assistant Registrar (P&S), IISER Mohali, on above given date and time. If the tender opening date happens to be on a holiday or non-working day due to any other valid reason, the tender opening process will be attended on the next working day at same time and place. IISER Mohali will not be responsible for any error like missing of schedule data while downloading by the Bidder.
- III. The bidder shall upload the tender documents duly filled in and stamped by the authorized signatory on each and every page. Tender not submitted/uploaded in the prescribed form and as per the tender terms and conditions shall be liable for rejection.
- IV. The bidder shall upload scanned copy of the PAN Card, GST number VAT return, Service tax registration number duly signed and stamped.
- V. E-procurement system ensures locking on the scheduled date and time. The system will not accept any bid after the scheduled date and time of submission of bid.

INSTRUCTIONS

1. The Online bids should be submitted directly by the original manufacturer/supplier, If quotation is submitted/filled by any representative/agent/dealer then they must upload a authority certificate from the principal company for quoting the price otherwise such quotation will be rejected.
2. The quantity mentioned in this inquiry is and shall be deemed to be only approximate and will not in any manner be binding on the Institute. Before the deadline for submission of the online bid, IISER Mohali reserves the right to modify the tender document terms and conditions. Such amendment/modification will be notified on website against said tender ID.
3. The rates offered should be FOR Chandigarh/Mohali in case of firms situated outside Chandigarh/Mohali, and free delivery at the Institute premises in case of local firms. Supplier from outside India should mention the Ex-works/FOB/FCA/CIF/CIP price clearly. Conditional tenders will be summarily rejected.
4. In case of Ex-godown terms the amount of packaging forwarding freight etc. should clearly be indicated by percentage or lump sum amount. Institute has policy not to make any advance payments towards any purchase, Letter of credit can be opened if required.
5. THE INSTITUTE IS EXEMPTED FROM EXCISE AND CUSTOM DUTY under notification no- TU/V/RG-CDE(1062)1 201–CUSTOM DATED 30.08.2016.
6. Tax: This Institute is not exempted from the payment of GST. The current rate (i.e. percentage of GST should be clearly indicated included or excluded) wherever chargeable. Please also provide/upload the copy of PAN card, GST number duly self-attested.
7. The delivery period should be specifically stated. Earlier delivery will be preferred.
8. The firms are requested to provide/upload detailed description and specifications together with the detailed drawings, printed leaflets and literature of the article quoted. The name of the manufactures and country of manufacture should also invariably be stated. In the absence of these particulars, the quotation is liable for rejection.
9. Validity of offer: 90 days. The warranty period after satisfactory installation should be mentioned and firm should replace all manufacturing defect parts/ whole item under warranty without any extra cost including clearance, freight, taxes. Security deposit/ Bank Performance Guarantee @ 10 % of the value of supply order as per norms may be sought from the firms.
10. The right to reject all or any of the quotation and to split up the requirements or relax any or all the above conditions without assigning any reason is reserved by the IISER Mohali. For any corrigendum and addendum please be checked the website <https://eprocure.gov.in/eprocure/app> and <http://www.iisermohali.ac.in>
11. Disputes, if any, shall be subject to jurisdiction in the court of Mohali only.

–sd/-

(Mukesh Kumar)
Assistant Registrar (S&P)