

भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान मोहाली

मानव संसाधन विकास मंत्रालय, भारत सरकार द्वारा स्थापित सैक्टर 81,नॉलेज सिटी,प॰ओ॰ मनोली, एस॰ ए॰ एस॰ नगर,मोहाली, पंजाब 140306 INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH MOHALI (Established by Ministry of Human Resource Development, Govt. of India)

Sector-81, Knowledge city, PO-Manauli, SAS Nagar Mohali-140306, Punjab PAN No. - AAAAI1781K GST: 03AAAAI1781K2ZS

Phone : +91-172-2240086 & 2240121 • Fax : +91-172-2240124, 2240266 • http://www.iisermohali.ac.in • Email: stores@iisermohali.ac.in

CPPP/Institute Website

IISERM (1113)18/19-Pur

Dated : 13.12.2018

NOTICE INVITING E-TENDER

Online tenders are invited on behalf of Director, IISER Mohali in <u>TWO BID SYSTEM</u> for **SUPPLY AND INSTALLATION OF SUPER-RESOLUTION MICROSCOPY FACILITY AT IISER MOHALI** from reputed companies/firms/ individuals/ societies etc. those are in the similar business at CPPP i.e. https://eprocure.gov.in/eprocure/app. Tender documents may please be downloaded from the E-procurement portal website https://eprocure.gov.in/eprocure/app& Institute website www.iisermohali.ac.in.

-sd-(Mukesh Kumar) Assistant Registrar (S&P)



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान मोहाली

मानव संसाधन विकास मंत्रालय, भारत सरकार द्वारा स्थापित

सैक्टर 81.नॉलेज सिटी.प॰ ओ॰ मनोली. एस॰ ए॰ एस॰ नगर.मोहाली. पंजाब 140306

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH MOHALI

(Established by Ministry of Human Resource Development, Govt. of India)

Sector-81, Knowledge city, PO-Manauli, SAS Nagar Mohali-140306, Punjab

PAN No. - AAAAI1781K GST: 03AAAAI1781K2ZS

Phone : +91-172-2240086 & 2240121 • Fax : +91-172-2240124, 2240266 • http://www.iisermohali.ac.in • Email: <u>stores@iisermohali.ac.in</u>

CPPP/Institute Website

E-TENDER NOTICE

Tender Ref IISERM(1113)18/19-Pur	Dated :- 13 th December, 2018
----------------------------------	--

Critical Date Sections

Sr.	Description	Date	Time
1.	Tender Publishing Date and time	13 th December, 2018	6:00pm
2.	Tender Document download start Date & Time	13 th December, 2018	6:00pm
3.	Bid Submission start Date & Time	13 th December, 2018	6:00pm
4.	Bid Submission End date and Time	07 th January, 2019	Upto 11:00am
5.	Tender opening Date and Time	08 th January, 2019	At 11.30 am

Online tenders are invited on behalf of the Director, IISER Mohali in **TWO BID SYSTEM** from the original manufacturer/supplier at CPPP i.e. for following item(s) https://eprocure.gov.in/eprocure/app. Tender documents may please be downloaded from the E-procurement portal website https://eprocure.gov.in/eprocure/app & Institute website www.iisermohali.ac.in.Tender fee in shape of DD/Banker Cheque of Rs 590/- (Nonrefundable) and EMD of Rs.7,00,000/- should be submitted by DD /Banker Cheque/FDR/ Bank Guarantee in favour of the Registrar, IISER Mohali payable at Mohali. However, scanned copy of the both tender fee and EMD should be upload on website along with technical bid part. The hard copy of the same in original to be send to the address mentioned below duly superscribing the supply/work name and reference/ tender ID on the envelope and same must reach before opening the bid and if not received within due date the bid will be rejected summarily.

The Original EMD and Tender Fee should be sent to:

Assistant Registrar (S&P) Indian Institute of Science Education and Research, Mohali Sector 81, SAS Nagar, Mohali, Punjab, India, Pin: 140306

Non-receipt of original EMD and Tender Fee will lead to rejection of tender.

Item Details:				
S. No.	Details of Specifications of the Stores	Qt y		
1.	SUPPLY AND INSTALLATION OF SUPER-MICROSCOPE FACILITY AT IISER MOHALI	01 No.		
	SUPER-RESOLUTION MICROSCOPY FACILITY AT IISER MOHALI			
	SUPER-RESOLUTION IMAGING USING SINGLE MOLECULE LOCALIZATION SYSTEM WITH HIGH-RESOLUTION AND SENSITIVE SPECTRAL CONFOCAL WORKSTATION			
	A super-resolution imaging workstation should include high-resolution and sensitive spectral confocal imaging for fixed and live sample imaging for cell biology/developmental biology applications, protein-protein interactions and imaging of modal organisms. The system should be able to perform single-molecule localization using stochastic optical reconstruction microscopy (STORM) / photo-activated localization microscopy (PALM) and should include multichannel fluorescence imaging with Z-stack, time-lapse including co-localization, fluorescence resonance energy transfer (FRET), fluorescence recovery after photobleaching (FRAP), photo-activation and conversion imaging and analysis. Apart from conventional confocal system the system should be equipped for carrying out quick high-resolution (\leq 150 nm) imaging that will be needed prior to more involved super-resolution single-molecule localization. The illumination source(s) would be primarily laser-based across the visible spectrum with spectral detection and should be a direct outcome of the microscope and not a mere result of any post-processing. The system should be available with the configuration mentioned below:			
	A. Motorized Inverted Fluorescence Research Microscope:			
	 a) Fully Motorized Inverted Fluorescence Research Microscope for BF/DIC/phase contrast/Fluorescence preferably with dedicated touch screen TFT display for controlling motorized components of the microscope. b) Programmable motorized X-Y scanning stage, Universal sample holders for slides, 35/60 mm Petri dish, labtek chambers with multipoint, tile and mosaic imaging software. 			
	c) A fast piezo focusing stage insert for fast z stack imaging with travel range of 100 microns or better.			
	d) IR LED (840nm Hardware based) focus drift compensation mechanism for long term live cell imaging application should be available as standard with the system and controlled by the software.			
	 e) 12V/100W halogen illumination for transmitted light & 120W metal halide illumination for Fluorescence should be offered. 			
	f) Motorized 6 position DIC nosepiece, Universal Motorized Condenser NA 0.55 or better with modules for DIC, 6 position fluorescence turrets for accommodating fluorescent filters for			

- sample visualization and camera-based imaging.
- g) High precision Z-focus drive with step size of 15 nm or better.
- h) High resolution confocal grade objectives of 10X/0.4 or better, 20X/0.80, 40x/1.30oil, 60/63x/1.40oil immersion or better for Imaging and SR work. Dedicated TIRF objective 100X/1.46 or better for TIRF and single molecule localization should be offered. Additionally, water immersion 40X/60X (1.20) WD 0.20 mm or better should also be quoted for FCS/FCCS

work in cells and solutions.

- i) Automated shift free DIC accessories for all objectives.
- j) Band pass fluorescent filters for DAPI, GFP and Cy3 should be offered.
- k) An active anti-vibration table with compressed air damping, bread board table top with M-6 threading for the complete microscope system.
- I) Monochrome cooled CCD camera, 2/3"Chip with 2.6 million or better net effective pixel resolution (FireWire based/USB III) with frame rate of 30 fps or better at full format, controlled by the same confocal software for multichannel, z stack, time lapse wide field imaging. Deconvolution software
- m) Facility for live cell imaging including Incubation system with Temperature, CO2, humidity control and complete safety regulations should be offered. The parameters for Incubation system should be controlled by confocal software.
- n) Dedicated attachment for converting Inverted Microscope to Upright Microscope for Tissue Imaging with depth up to 200 microns or better (depending on transparency of samples). Additionally, one Water Dipping 25X and 60X, NA 1.0 and WD of 2.1mm or better for deep tissue imaging should be incorporated.

B. Spectral confocal imaging unit with built-in highly sensitive detectors:

- a) Laser point scanning and Confocal detection unit with built-in Spectral PMT and HyD/GaAsP Spectral detectors. All detectors should be capable of working in Intensity and Spectral mode Imaging. Should be capable of simultaneous detection and separation of minimum 4 fluorophores or more based on high sensitive GaAsP/HyD or equivalent detectors with QE 45% or more.
- b) Scanner unit should have laser ports for Vis, UV and IR lasers. It should include high efficient excitation laser suppression beam splitting device with low angle of incidence dichorics.
- c) The scanner should have real "ROI" scan capability for fast scan. Maximum scan resolution should be at least 4K x 4K or better per channel and should reduce to 16X16 resolution.
- d) Scan speed should be 12 15 fps or better @ 512x512. The scan head should be able to perform fast dynamic live cell time lapse imaging with a high speed of at least 200 fps or better @512X 32 resolution.
- e) Transmitted PMT for laser based DIC imaging should be included.
- f) The scan field diagonal should be 18 mm or better. Scan Zoom range 1X to 40X with increments of 0.1X.

C. Laser module with AOTF control:

- a) Multi-line Ar laser with 458/488/514nm.
- b) DPSS 561nm
- c) HeNe 594 nm
- d) HeNe 633 nm
- e) Blue diode (UV) Laser 405/408 nm

All visible & UV lasers should be connected to the scan head through fiber optic cable and should be controlled through AOTF for fast laser switching and attenuation in pixel precise synchronization with the laser scanner for Real ROI scan for FRAP, Photo activation/conversion experiments. All the laser lines should be controlled through a computerized AOTF device for fast laser switching and attenuation.

D. FCS/FCCS system for single molecule detection: Based on minimum 2 channel FCS/FCCS

with high sensitivity and minimum after pulsing. The FCS unit should perform auto and cross correlation measurements in live cells and solution for a wide range of dyes and proteins. The unit should have the facility for elimination/suppression of other excitation laser lines. All laser lines for confocal imaging should be capable of working in FCS/FCCS mode. FCS measurement software for auto- and cross-correlation capabilities should be quoted. System should also be included with POL Anisotropy accessories to perform quantitative anisotropy experiments and should be duly supported by literatures and references.

E. Real-time Online High-Resolution Imaging:

- a) Fully automated, real-time and Online SR attachment with suitable sensitive detectors for complete visible spectrum.
- b) Should be able to achieve Lateral resolution of 120-150 nm and axial resolution of 350 450nm.
- c) The system should be capable of working with cell culture, cell lines, modal organisms and tissues with depth of penetration of 100 microns or more.

F. Single Molecule Localization System (STORM / PALM) (for Proteins, organic dyes and fluorophores)

- a) STORM / PALM for single molecule localization to achieve resolution down to 20 nm or better in XY and 130 nm in Z (TIRF regime). Should be supported for all laser lines in the spectral region.
- b) Acquisition in various modes i.e. localization, single particle tracking, fiducial correction, activation power controlled, Wide field, TIRF imaging, HILO mode and ultra-high TIFR modes should be available.
- c) On-line processing for on-the-fly PALM/STORM analysis. Data should become available as the images are being acquired.
- d) Image based auto focussing and activation power ramp. Frame time in tracking down to 40 ms. Multi-channel or multi-color acquisition in a sequential mode should be possible for applications in Cytoskeleton, Focal adhesions, Membrane organization, Vesicle transport and organelle architecture.
- e) Pixel-by-pixel overlap calibration and sample independent drift compensation by fiducial tracking on all detection channels.
- f) Dedicated high sensitivity EMCCD camera for single molecule imaging with 512X512 pixel resolution and QE of 95% with frame rate of 55 fps or better at full format, TE cooling of -100 °C, 16 μm X 16 μm pixel size for best sensitivity coupled to the camera port of the microscope.

g) Laser module with AOTF control for Single Molecule Localization System:

High-power solid-state lasers 488 nm (100 mW or better for Ax 488, GFP, FITC fluorophores), 561 nm (100 mW or better for TRITC, Rhodamine, Texas Red, Cy3, Ax 546), 405 nm (50 mW or better for DAPI, Hoechst dyes), 640 nm (100mW or better for Ax 647, Cy5, Dil, DIO). All the laser lines should be controlled through a computerized AOTF device for fast laser switching and attenuation.

G. Control computer and Monitor:

Latest 64-bit control computer with Intel Xeon 6 Core Processor, DDR RAM 64 GB HDD: 4 TB SATA upgradable to 8 TB or better, DVD, Super Multi SATA +R/RW, Graphics: AT Fire GL V5200 256MB DH DVI, Gigabit Ethernet, Win 7 Ultimate 64 bit, USB 2.0, Fire wire. Large 32" LCD/ TFT monitor.

System control and Imaging Software:

- a) Software should be capable of controlling Motorized components of microscope, digital camera, confocal scan head, laser control including AOTF and Image acquisition & processing for confocal and super resolution imaging.
- b) Saving of all system parameters with the image for repeatable/reproducible imaging.
- c) Line, curved line, frame, Z-stack, Time series imaging capabilities.
- d) Real ROI bleach for FRAP, Photo-activation/conversion experiments.
- e) FRET imaging as well as Quantitative data analysis capability.
- f) Standard geometry measurements like length, areas, angles etc including intensity measurements.
- g) Advanced 3D image reconstruction with rendering from a Z-stack image series.
- h) Co-localization and histogram analysis with individual parameters.
- i) Spectral un-mixing with fingerprinting for separation of overlapping excitation/emission spectra of fluorophores.
- j) Image acquisition and processing tolls for SR with various modes of visualization/ analysis tools should be available.
- k) Additional Offline software with complete features as the main software with high end dedicated PC and monitor (same specs as main PC) should be available.

Note:

- 1. Bidders should clearly specify the after sales service/application support capabilities without any additional cost.
- 2. Warranty for the complete system and additional 4 years AMC should be included.
- 3. Provide all information as regards pre-installation requirements (i.e. room, environment) for system installation.
- 4. Online UPS for the complete system including lasers should be included in the supply. The system should have a dedicated online branded UPS system with at least 30 min back up for the whole system.
- 5. Detailed list of users of the system in India with contact details to be provided.
- 6. Onsite training should be available.

NB :-

- I. The online updated Price BOQ is in INR format. If bidder want to quote other than INR please specify the quoted currency in the technical bid/part and fill the amount in same updated BOQ.
- II. Please bifurcate the price on shipping terms i, e, Ex-works -> FCA/FOB -> CIP/CIF in price BOQ and specify the same in technical bid without price.

SUBMISSION OF TENDER

I. All bid/ tender documents are to be uploaded online at Central Public Procurement portal i.e. <u>https://eprocure.gov.in/eprocure/app</u> only and in the designated cover/ part on the website against tender ID. Tenders/ bids shall be accepted only through online

mode and no manual submission of the same shall be entertained except tender fee and EMD. Late tenders will not be accepted.

- II. The online bids shall be opened at the office of the Assistant Registrar (P&S), IISER Mohali, on above given date and time. If the tender opening date happens to be on a holiday or non-working day due to any other valid reason, the tender opening process will be attended on the next working day at same time and place. IISER Mohali will not be responsible for any error like missing of schedule data while downloading by the Bidder.
- III. The bidder shall upload the tender documents duly filled in and stamped by the authorized signatory on each and every page. Tender not submitted/uploaded in the prescribed form and as per the tender terms and conditions shall be liable for rejection.
- IV. The bidder shall upload scanned copy of the PAN Card, GST number duly signed and stamped.
- V. E-procurement system ensures locking on the scheduled date and time. The system will not accept any bid after the scheduled date and time of submission of bid.

INSTRUCTIONS

- 1. The Online bids should be submitted directly by the original manufacturer/supplier, If quotation is submitted/filled by any representative/agent/dealer then they must upload a authority certificate from the principal company for quoting the price otherwise such quotation will be rejected.
- 2. The quantity mentioned in this inquiry is and shall be deemed to be only approximate and will not in any manner be binding on the Institute. Before the deadline for submission of the online bid, IISER Mohali reserves the right to modify the tender document terms and conditions. Such amendment/modification will be notified on website against said tender ID.
- 3. The rates offered should be FOR Chandigarh/Mohali in case of firms situated outside Chandigarh/Mohali, and free delivery at the Institute premises in case of local firms. Supplier from outside India should mention the Ex-works/FOB/FCA/CIF/CIP price clearly. Conditional tenders will be summarily rejected.
- 4. In case of Ex-godown terms the amount of packaging forwarding freight etc. should clearly be indicated by percentage or lump sum amount. Institute has policy not to make any advance payments towards any purchase, Letter of credit can be opened if required.
- 5. THE INSTITUTE IS EXEMPTED FROM EXCISE AND CUSTOM DUTY under notification no- TU/V/RG/-CDE(1062)/201 CUSTOM DT.30.08.2016.
- 6. Tax: This Institute is not exempted from the payment of GST. The current rate (i.e. percentage of GST should be clearly indicated included or excluded) wherever chargeable. Please also provide/upload the copy of PAN card, GST number duly self-attested.
- 7. Concessional GST is applicable for all the items purchased for Research labs vide Ministry of Finance, notification no. 45/22017 dated 14.11.2017 and 47/2017 dated

14.11.2017.

- 8. Bidder/s quoting in currency other than **Indian Rupee** (**INR**) should explicitly mention the currency in which tender quoted wherever applicable in <u>Technical Bid</u> along the tender documents.
- 9. The delivery period should be specifically stated. Earlier delivery will be preferred.
- 10. The firms are requested to provide/upload detailed description and specifications together with the detailed drawings, printed leaflets and literature of the article quoted. The name of the manufactures and country of manufacture should also invariably be stated. In the absence of these particulars, the quotation is liable for rejection.
- 11. Validity of offer: 90 days. The warranty period after satisfactory installation should be mentioned and firm should replace all manufacturing defect parts/ whole item under warranty without any extra cost including clearance, freight, taxes. Security deposit/ Bank Performance Guarantee @ 10 % of the value of supply order as per norms may be sought from the firms.
- 12. The right to reject all or any of the quotation and to split up the requirements or relax any or all the above conditions without assigning any reason is reserved by the IISER Mohali. For any corrigendum and addendum please be checked the website <u>https://eprocure.gov.in/eprocure/app</u> and <u>http://www.iisermohali.ac.in</u>
- 13. Disputes, if any, shall be subject to jurisdiction in the court of Mohali only.

-sd/-(Mukesh Kumar) Assistant Registrar (S&P)