



# भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान मोहाली

मानव संसाधन विकास मंत्रालय, भारत सरकार द्वारा स्थापित

सैक्टर-81, नॉलेज सिटी, प० ओ० मनोली, एस० ए० एस० नगर, मोहाली, पंजाब 140306

**INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH MOHALI**

(Established by Ministry of Human Resource Development, Govt. of India)

Sector-81, Knowledge city, PO-Manauli, SAS Nagar Mohali-140306, Punjab

PAN No. - AAAAI1781K GSTIN No:- 03AAAAI1781K2ZS

• Phone : +91-172-2240086 & 2240121 • Fax : +91-172-2240124, 2240266 • <http://www.iisermohali.ac.in> • Email: [stores@iisermohali.ac.in](mailto:stores@iisermohali.ac.in)

CPPP/Institute Website

IISERM (1244)19/20Pur

Dated: 16<sup>th</sup> August 2019

## **NOTICE INVITING E-TENDER**

Online tenders are invited on behalf of Director, IISER Mohali in **SINGLE BID SYSTEM** for the **Supply and installation of 4° C Glass Door Refrigerator and -20° Solid Door Refrigerator** as per technical specification and details given below and BOQ list from the original printing houses/firms at CPPP i. e. <https://eprocure.gov.in/eprocure/app>. Tender documents may please be downloaded from the E-procurement portal website <https://eprocure.gov.in/eprocure/app> & Institute website [www.iisermohali.ac.in](http://www.iisermohali.ac.in).

-sd-

(Mukesh Kumar)  
Assistant Registrar (S&P)



# भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान मोहाली

मानव संसाधन विकास मंत्रालय, भारत सरकार द्वारा स्थापित

सैक्टर-81, नॉलेज सिटी, प० ओ० मनोली, एस० ए० एस० नगर, मोहाली, पंजाब 140306

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH MOHALI

(Established by Ministry of Human Resource Development, Govt. of India)

Sector-81, Knowledge city, PO-Manauli, SAS Nagar Mohali-140306, Punjab

PAN No. - AAAAII781K GSTIN No:- 03AAAAII781K2ZS

• Phone : +91-172-2240086 & 2240121 • Fax : +91-172-2240124, 2240266 • <http://www.iisermohali.ac.in> • Email: [stores@iisermohali.ac.in](mailto:stores@iisermohali.ac.in)

## **E-TENDER NOTICE**

|                                   |                                       |
|-----------------------------------|---------------------------------------|
| Tender Ref.- IISERM(1244)19/20Pur | Dated :- 16 <sup>th</sup> August 2019 |
|-----------------------------------|---------------------------------------|

### **Critical Date Sections**

| Sr. | Description                                | Date                         | Time         |
|-----|--|------------------------------|--------------|
| 1.  | Tender Publishing Date and time            | 16 <sup>th</sup> August 2019 | 6:00pm       |
| 2.  | Tender Document download start Date & Time | 16 <sup>th</sup> August 2019 | 6:00pm       |
| 3.  | Bid Submission start Date & Time           | 16 <sup>th</sup> August 2019 | 6:00pm       |
| 4.  | Bid Submission End date and Time           | 05 <sup>th</sup> Sept 2019   | Upto 11:00am |
| 5.  | Tender opening Date and Time               | 06 <sup>th</sup> Sept 2019   | At 11:30am   |

Online tenders are invited on behalf of Director, IISER Mohali in **SINGLE BID SYSTEM** {Technical and commercial separately} for the following item(s) from the original manufacturer/supplier at CPPP i.e. <https://eprocure.gov.in/eprocure/app>. Tender documents may please be downloaded from the E-procurement portal website <https://eprocure.gov.in/eprocure/app> & Institute website [www.iisermohali.ac.in](http://www.iisermohali.ac.in). Tender fee in shape of DD/Banker Cheque of Rs 590/- (Non-refundable) and EMD of Rs 5000/- should be submitted by DD /Banker Cheque/FDR/ Bank Guarantee in favour of the Registrar, IISER Mohali payable at Mohali. However, scanned copy of the both tender fee and EMD should be upload on website along with technical bid part. The hard copy of the same in original to be send to the address mentioned below duly superscribing the supply/work name and reference/ tender ID on the envelope and same must reach before opening the bid and if not received within due date the bid will be rejected summerly.

### **The Original EMD and Tender Fee should be sent to:**

Assistant Registrar (S&P)  
Indian Institute of Science Education and Research Mohali  
Sector-81, Knowledge City, SAS Nagar, Mohali, Punjab,  
India, Pin: 140306

**Non-receipt of original EMD and Tender Fee will lead to rejection of tender.**

**Item Details: -**

| <b>Sr.</b> | <b>Description</b>  | <b>Qty<br/>(in<br/>units)</b> |
|------------|---|-------------------------------|
| <b>1.</b>  | <b>Supply and installation of 4° C Glass Door Refrigerator</b><br><br><b><u>Technical Specifications</u></b><br><br><ol style="list-style-type: none"><li>1. The system should be a refrigerator with at least two swing out glass doors.</li><li>2. The doors should have an in-built locking mechanism.</li><li>3. The capacity of the usable area should be at least 950 liters.</li><li>4. The system should have dimensions of at least 45 X 28 X 82 inches but dimensions should be no more than 50 X 30 X 85 inches.</li><li>5. The system should have an adjustable temperature range (adjustable through an in-built digital temperature controller) of at least 2°C to 8°C and the temperature should be of the storage/usable area.</li><li>6. The temperature of the storage/usable area of the refrigerator should be displayed in an in-built digital temperature display.</li><li>7. The error in temperature should not be more than +/- 1°C and the system should be calibrated by the vendor.</li><li>8. The system should be able to maintain uniform temperature in the entire in-built storage/usable area.</li><li>9. The system should have at least two covered ports (one on each side along the width) such that wires can be passed from the interior (storage area) of the refrigerator to the outside.</li><li>10. The ports mentioned in specification 8 should not influence the temperature of the in-built storage area.</li><li>11. The system should be able to accommodate at least 8 adjustable metallic shelves (either of stainless steel or painted).</li><li>12. The system should be supplied with the shelves describe in specification 10.</li><li>13. The system should not use CFC.</li><li>14. The refrigerator should have interior lights.</li><li>15. The system should be supplied with plugs compatible with the Indian power grid voltage and frequency (230V and 50Hz).</li><li>16. The system should conform to the standards set by the Bureau of Indian Standards (BIS) and the vendor should provide documents to prove the same.</li><li>17. The system should be supplied with a voltage stabilizer for High or Low voltage cut out with timer delay.</li><li>18. The vendor/manufacturer should have at least 10 installations of the system (or similar systems) in Chandigarh area and the vendor/manufacturer should furnish documents to support this claim.</li><li>19. The system should not be an end of life model.</li><li>20. The system should come with two years of comprehensive warranty.</li><li>21. Compliance sheet with yes/no is required for tender specifications.</li></ol> | <b>0</b>                      |
| <b>2.</b>  | <b>Supply and installation of -20° Solid Door Refrigerator</b><br><br><b><u>Technical Specifications</u></b><br><br><ol style="list-style-type: none"><li>1. The system should be a deep freezer with a single solid door.</li></ol>  | <b>0</b>                      |

2. The door should have an in-built locking mechanism.
3. The capacity of the usable area should be at least 300 liters.
4. The system should have an adjustable (through a digital temperature controller) temperature range of at least -10°C to -21°C and the temperature should be of the usable/storage area.
5. The error in temperature should not be more than +/- 1°C and the system should be calibrated by the vendor.
6. The temperature of the storage/usable area of the freezer should be displayed in a digital temperature display.
7. The system should be able to maintain uniform temperature in the entire storage/usable area.
8. The system should be able to accommodate at least 3 glass shelves and at least 5 pull out draws.
9. The system should be supplied with the shelves and draws describe in specification 7.
10. The system should not use CFC.
11. The system should not automatically defrost.
12. The freezer should have interior lights.
13. The system should be supplied with plugs compatible with the Indian power grid voltage and frequency.
14. The system should conform to the standards set by the Bureau of Indian Standards (BIS) and the vendor should provide documents to prove the same.
15. The system should be supplied with a voltage stabilizer for High or Low voltage cut out with timer delay.
16. The vendor/manufacturer should have at least 10 installations of the system (or similar systems) in Chandigarh area and the vendor/manufacturer should furnish documents to support this claim.
17. The system should not be an end of life model.
18. The system should come with three years of comprehensive warranty.
19. Compliance sheet with yes/no is required for tender specifications.

**NB:-**

- I. The online updated Price BOQ is in INR format. If bidder want to quote other than INR please specify the quoted currency in the technical bid/part and fill the amount in same updated BOQ.
- II. Please bifurcate the price on shipping terms i, e, Ex-works -> FCA/FOB -> CIP/CIF in price BOQ and specify the same in technical bid without price.
- III. **The Online bids should be submitted directly by the original manufacturer/supplier, If quotation is submitted/filled by any representative/agent/dealer then they must upload a authority certificate from the principal company.**
- IV. **Kindly specify the weight of the product including weight of packing for assessment of Freight charges if quoted in foreign currency.**

**SUBMISSION OF TENDER**

- I. All bid/ tender documents are to be uploaded online at Central Public Procurement portal i.e. <https://eprocure.gov.in/eprocure/app> only and in the designated cover/ part on the website against tender ID. Tenders/ bids shall be accepted only through online mode and no manual submission of the same shall be entertained except tender fee and EMD. Late tenders will not be accepted.
- II. The online bids shall be opened at the office of the Assistant Registrar (P&S), IISER Mohali, on above given date and time. If the tender opening date happens to be on a holiday or non-working day due to any other valid reason, the tender opening process will be attended on the next working day at same time and place. IISER Mohali will not be responsible for any error like missing of schedule data while downloading by the Bidder.
- III. The bidder shall upload the tender documents duly filled in and stamped by the authorized signatory on each and every page. Tender not submitted/uploaded in the prescribed form and as per the tender terms and conditions shall be liable for rejection.

- IV. The bidder shall upload scanned copy of the PAN Card, GST number duly signed and stamped. **Also bidders applying against ‘MSME/NSIC Certificate’ issued by appropriate Authority, should ensure that the certificate attached is relevant to the area of service/supply. For example, If the tender is for “supply & installation of Desktop” the certificate should be issued for activity/area of “Computer supply and services activities etc” otherwise bid will be REJECTED without notice.**
- V. E-procurement system ensures locking on the scheduled date and time. The system will not accept any bid after the scheduled date and time of submission of bid.

## **INSTRUCTIONS**

1. The Online bids should be submitted directly by the original manufacturer/supplier, If quotation is submitted/filled by any representative/agent/dealer then they must upload a authority certificate from the principal company for quoting the price otherwise such quotation will be rejected.
2. The quantity mentioned in this inquiry is and shall be deemed to be only approximate and will not in any manner be binding on the Institute. Before the deadline for submission of the online bid, IISER Mohali reserves the right to modify the tender document terms and conditions. Such amendment/modification will be notified on website against said tender ID.
3. The rates offered should be FOR Chandigarh/Mohali in case of firms situated outside Chandigarh/Mohali, and free delivery at the Institute premises in case of local firms. Supplier from outside India should mention the Ex-works/FOB/FCA/CIF/CIP price clearly. Conditional tenders will be summarily rejected.
4. In case of Ex-godown terms the amount of packaging forwarding freight etc. should clearly be indicated by percentage or lump sum amount. Institute has policy not to make any advance payments towards any purchase, Letter of credit can be opened if required.
5. THE INSTITUTE IS EXEMPTED FROM EXCISE AND CUSTOM DUTY under notification no- TU/V/RG/-CDE(1062)/201 CUSTOM DT.30.08.2016.
6. Tax: This Institute is not exempted from the payment of GST. The current rate (i.e. percentage of GST should be clearly indicated included or excluded) wherever chargeable. Please also provide/upload the copy of PAN card, GST number duly self-attested.
7. Concessional GST is applicable for all the items purchased for Research labs vide Ministry of Finance, notification no. 45/22017 dated 14.11.2017 and 47/2017 dated 14.11.2017.
8. Bidder/s quoting in currency other than **Indian Rupee (INR)** should explicitly mention the currency in which tender quoted wherever applicable in Technical Bid along the tender documents.
9. The delivery period should be specifically stated. Earlier delivery will be preferred.
10. The firms are requested to provide/upload detailed description and specifications together with the detailed drawings, printed leaflets and literature of the article quoted. The name of the manufactures and country of manufacture should also invariably be stated. In the absence of these particulars, the quotation is liable for rejection.
11. Validity of offer: 90 days. The minimum warranty period will be one year where not specified in along with Technical specification after satisfactory installation should be mentioned and firm should replace all manufacturing defect parts/ whole item under warranty without any extra cost including clearance, freight, taxes. Security deposit/ Bank Performance Guarantee @ 10 % of the value of supply order as per norms may be sought from the firms.

12. The right to reject all or any of the quotation and to split up the requirements or relax any or all the above conditions without assigning any reason is reserved by the IISER Mohali. For any corrigendum and addendum please be checked the website <https://eprocure.gov.in/eprocure/app> and <http://www.iisermohali.ac.in>

13. Disputes, if any, shall be subject to jurisdiction in the court of Mohali only.

–sd–

(Mukesh Kumar)

Assistant Registrar (P&S)