



# भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान मोहाली

मानव संसाधन एवं विकास मंत्रालय, भारत सरकार द्वारा स्थापित  
सैक्टर 81, नॉलेज सिटी, पी. ओ. मनोली, एस. ए. एस. नगर, मोहाली, पंजाब 140306  
INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH MOHALI  
(Established by Ministry of Human Resource Development, Govt. of India)  
Sector-81, Knowledge city, PO-Manauli, SAS Nagar Mohali-140306, Punjab  
PAN No. - AAAAI1781K TAN No. PTLI10692D

• Phone : +91-172-2293118-19 • Fax : +91-172-2240124, 2240266 • <http://www.iisermohali.ac.in> • Email: [stores@iisermohali.ac.in](mailto:stores@iisermohali.ac.in)

E-mail/CPPP/Website

IISER M(475)14/15 Pur.

Date : 31<sup>st</sup> December 2014

## NOTICE INVITING QUOTATION

Sealed Quotations are invited on behalf of Director IISER Mohali for the purchase of following item so as to reach latest by 21<sup>st</sup> January 2015 before 1 pm. The Quotations will be opened on the same day at 4 pm in the presence of tenderers, if any :-

Sr.	Description	Qty
01.	<b>Vacuum Concentrator</b> Specifications : <ol style="list-style-type: none"><li>1. Required vacuum concentrator should concentrate samples like DNA, RNA and Proteins etc. with precise temperature, rpm and vacuum control through</li><li>2. Concentrator chamber should be acid resistant and Teflon coated.</li><li>3. The rotor to hold 40 or above for 1.5/2 ml tubes should be supplied along with the system and other size rotors should also be quoted as optional accessories.</li><li>4. Temperature range should be at least ambient to 80°C or with lower limit of 5°C.</li><li>5. Concentration of sample should be fully programmable for temperature, run time and vacuum level.</li><li>6. Concentrator rpm should be minimum 1500 or adjustable up to 2000 rpm.</li><li>7. Concentrator should be supplied with acid resistant cold trap made of 316 grade seamless stainless steel with external cooling coil with digital alpha numeric temperature display.</li><li>8. Cold trap should attain minimum temperature of -50°C or below and should be of at least 4 liter capacity.</li><li>9. System should be supplied complete along with chemical resistant vacuum pump of at least 37 liters/min displacement capacity and should attain minimum of 9 mbar ultimate vacuum.</li><li>10. Vendor should supply a cart suitable to stably hold and operate the system.</li><li>11. Complete concentrator should be upgradable to function as a lyophilizer with minimum 4 to 8 tree type manifold, and accessories for the same should also be quoted.</li><li>12. All tubing and connectors should be included as the part of concentrator.</li><li>13. All items should come from the same principle company.</li><li>14. Machine manufacture should compliant to CE, DS/EN and ISO certification standards and certificates should be enclosed.</li><li>15. The vendor should submit case specific authorization letter and certificate from the manufacturer for supply of spare parts for at least 10 years of purchase, if quoted by the distributor. If manufacturer is quoting directly, the company should submit a certificate committing to make available the spare parts for at least 10 years from purchase on their letterhead.</li><li>16. The vendor should give a warranty of at least 3 years from the date of installation of the system.</li></ol>	01

Sd/-  
(Mukesh Kumar)

## **INSTRUCTIONS**

1. Inquiry will be sent by UPC/Courier/Speed Post and IISER Mohali will not be liable for any kind of Postal delay.
2. The Quotation should be addressed to the Director IISER Mohali invariably giving on the envelop Reference No. with Due date & time.
3. One time importers from China with custom made specifications are highly discouraged.
4. The quantity mentioned in this inquiry is and shall be deemed to be only approximate and will not in any manner be binding on the Institute.
5. Firms will quote separately for each article.
6. The rates offered should be FOR Chandigarh/Mohali. In case of firms situated outside Chandigarh/Mohali, and free delivery at the Institute premises in case of local firms. Supplier outside India please mention the FOB/FCA price clearly.
7. In case of Ex-godown terms the amount of packaging forwarding freight etc should clearly be indicated by percentage or lump sum amount. Institute has policy not to make any advance payment towards any purchase, Letter of credit can be opened if required.
8. In case of foreign consignment CIF/FOB should be mention clearly. THE INSTITUTE IS EXEMPTED FROM EXCISE AND CUSTOM DUTY.
9. SALES TAX :- This Institute is not exempted from the payment of Sales Tax/Service Tax/VAT. The current rate (i.e. percentage of Sales Tax should be clearly indicated included or excluded) wherever chargeable.
10. The delivery period should be strictly adhered, the firm have to execute the supply according to purchase order.
11. The firms are requested to give detailed description and specifications together with the detailed drawings, printed leaflets and literature of the Article quoted. The name of the manufactures and country of manufacture should also invariably be stated. In the absence of these particulars the quotation is liable for rejection.
12. Validity of offer: 90 days.
13. Late or delayed quotation will not be accepted.
14. Service Facility: Please mention the nearest service centre to IISER Mohali and the time required to attend the call
15. The right to reject all or any of the quotation and to split up the requirements or relax any or all the above conditions without assigning any reason is reserved.