



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान मोहाली

मानव संसाधन एवं विकास मंत्रालय, भारत सरकार द्वारा स्थापित
सैक्टर 81, नॉलेज सिटी, पी. ओ. मनोली, एस. ए. एस. नगर, मोहाली, पंजाब 140306
INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH MOHALI
(Established by Ministry of Human Resource Development, Govt. of India)
Sector-81, Knowledge city, PO-Manauli, SAS Nagar Mohali-140306, Punjab
PAN No. - AAAAI1781K TAN No. PTLI10692D

• Phone : +91-172-2293118-19 • Fax : +91-172-2240124, 2240266 • <http://www.iisermohali.ac.in> • Email: stores@iisermohali.ac.in

E-mail/CPPP/Website

IISER M(479)14/15 Pur.

Date : 01st January 2015

NOTICE INVITING QUOTATION

Sealed Quotations are invited on behalf of Director IISER Mohali for the purchase of following item so as to reach latest by 22nd January 2015 before 1 pm. The Quotations will be opened on the same day at 4 pm in the presence of tenderers, if any :-

Sr.	Description	Qty
01.	Multi-position magnetic stirrer with five stirrer position and integrated temperature control panel. (Technical specifications attached)	01

Sd/-
(Mukesh Kumar)
Assistant Registrar (S&P)

INSTRUCTIONS

1. Inquiry will be sent by UPC/Courier/Speed Post and IISER Mohali will not be liable for any kind of postal delay.
2. The Quotation should be addressed to the Director IISER Mohali invariably giving on the envelop reference number, last date & time of receipt of tender and date & time of opening of the tender.
3. One time importers from China with custom made specifications are highly discouraged.
4. The quantity mentioned in this inquiry is and shall be deemed to be only approximate and will not in any manner be binding on the Institute.
5. Firms will quote separately for each article.
6. The rates offered should be FOR Chandigarh/Mohali, in case of firms situated outside Chandigarh/Mohali, and free delivery at the Institute premises in case of local firms. Suppliers outside India may please mention the FOB/FCA price clearly.
7. In case of Ex-godown terms the amount of packaging forwarding freight etc should clearly be indicated by percentage or lump sum amount. Institute has policy not to make any advance payment towards any purchase, Letter of credit can be opened if required.
8. In case of foreign consignment CIF/FOB should be mentioned clearly. THE INSTITUTE IS EXEMPTED FROM EXCISE AND CUSTOM DUTY.
9. SALES TAX :- This Institute is not exempted from the payment of Sales Tax/Service Tax/VAT. The current rate (i.e. percentage of Sales Tax should be clearly indicated included or excluded) wherever chargeable.
10. The delivery period should be strictly adhered, the firm have to execute the supply according to purchase order.
11. The firms are requested to give detailed description and specifications together with the detailed drawings, printed leaflets and literature of the Article quoted. The name of the manufactures and country of manufacture should also invariably be stated. In the absence of these particulars, the quotation is liable for rejection.
12. Validity of offer: 90 days.
13. Late or delayed quotation will not be accepted.
14. Service Facility: Please mention the nearest service centre to IISER Mohali and the time required to attend the call.
15. The right to reject all or any of the quotation and to split up the requirements or relax any or all the above conditions without assigning any reason is reserved.

Multi-position magnetic stirrer with five stirrer position and integrated temperature control plate

Specifications:

1. Precise temperature distribution on the heating plate for performing series of five experiments at the same time
2. Simultaneously operating stirrers
3. Absolute consistency over sample conditions for the individual samples
4. The equipment should include suitable ac adapter/plug connections
5. Should demonstrate the model, if it does not already exist in IISER, Mohali.
6. One year warranty should be included.
7. The following technical specifications should be available:

Number of stirring positions	5
Stirring quantity max. per stirring position (H ₂ O)	0.4 l
Stirring position distance	80 - 100 mm
Speed deviation of individual stirring positions	≤5 %
Stirring quantity max. (H ₂ O)	2 l
Motor rating input	Approx. 7 W
Motor rating output	Approx. 2 W
Speed range	0 - 1100 rpm
Stirring bar length max.	Approx. 25 mm
Heat output	Approx. 175 W
Heating temperature range	Room temperature - 120 °C
Temperature medium max. (dep. on vessel)	Approx. 70 °C
Heat control	Approx. scale 1-10
Speed control	Approx. scale 1-10
Temperature constancy in medium	±2 K
Set-up plate material	Silicone
Dimensions (W x H x D)	Not exceeding 150 x 75 x 600 mm
Permissible ambient temperature	5 - 40 °C
Permissible relative humidity	80 %
Voltage	230 / 115 / 100 V
Frequency	50/60 Hz
Power input	185 W