



IISER Mohali

भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान मोहाली

मानव संसाधन एवं विकास मंत्रालय, भारत सरकार द्वारा स्थापित
सैक्टर 81, नॉलेज सिटी, पी. ओ. मनोली, एस. ए. एस. नगर, मोहाली, पंजाब –140306
INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH MOHALI

(Established by Ministry of Human Resource Development, Govt. of India)
Sector-81, Knowledge city, PO-Manauli, SAS Nagar Mohali-140306, Punjab

PAN NO. - AAAAI1781K TAN NO. PTLI10692D

• Phone : +91-172-2293118-19 • Fax : +91-172-2240124, 2240266 • <http://www.iisermohali.ac.in> • Email: stores@iisermohali.ac.in

IISERM(532) 15/16 Pur

E-mail/CPPP/Website

29th April 2015

Corrigendum

Sub: Extension of Due date for the submission the tender alongwith revised specification.

Dear Sir,

I am to inform you that the due date for the submission of Quotation is extended upto **08th May 2015** with modified specification (Add Qty. as below), Now you are requested to send your offer so as reach us within due date upto 1pm. The Quotation will be opened on the same day at 4pm in the presence of tenderers, if any.

-sd-
Assistant Registrar (S&P)

Semi-rigid cables of the following materials are needed

1. All cables must be semi-rigid with high density PTFE insulation and must have 50 ohms characteristic impedance.
 2. **~2.19 mm or ~ 0.085 inch Outer dia with 50 Ohm characteristic impedance.**
 3. Please specify minimum bend radius .
 4. Specify insertion loss per m or feet at few representative points like ~ 0.1GHz 1GHz 10 GHz or provide data sheets.
 5. All pieces must be 1m or 5 feet long whichever size is readily available.
 6. Quote the closet matching lengths in m or feet.
 7. **Please give price break for 15 feet, 30 feet ,45 feet, 60 feet & 90 feet or closest in meters for all items except g & h.**
 8. If MOQ is much larger give pricing for MOQ.
- a) Stainless steel inner and outer conductor
 - b) Stainless steel outer conductor and Silver plated stainless steel inner conductor
 - c)) Stainless steel outer conductor and Silver plated BeCu inner conductor
 - d)) Cupronickel outer conductor and Silver plated or Au plated CuNi inner conductor.
 - e)) Stainless steel outer conductor with 5 micron Cu plating and Silver plated brass inner conductor
 - f) **Copper outer conductor with tin** plating and silver plated copper inner conductor. quote low density PTFE if losses and phase distortions are smaller for this cable alone
 - g) Nb inner and outer conductor . Please quote for 0.5m 2 pieces.
 - h) NbTi Inner and outer conductor Please quote for 0.5 m 2 pieces

Please send quotes by e-mail to : stores@iisermohali.ac.in